

ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHHEISCHE REPUBLIK
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÉQUE
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod lisem 129/2, Troja, 182 00 Praha 8

CERTIFICATE

No.: 1250487

Product: Series of transformers for medical power supply JUVR-MED ECO line

Type: JOC Uxxxx-yyyy E
(xxxx=core construction, yyyy=number of winding rule)

Rating: Input: 230 or 400 V, 50/60 Hz, Safety class I
Output: 230 V, 2,5 kVA to 10 kVA

Ordering firm: ELEKTROKOV,a.s.ZNOJMO
Kotkova 3582/19, 669 02 Znojmo, Czech Republic

Manufacturer: ELEKTROKOV,a.s.ZNOJMO
Kotkova 3582/19, 669 02 Znojmo, Czech Republic

Factories: ELEKTROKOV,a.s.ZNOJMO
Kotkova 3582/19, 669 02 Znojmo, Czech Republic

Trade mark:

The test results are stated in the test-report No.: 252783-01/01 of: 09.12.2025

A sample of the product was found to be in conformity with:
EN IEC 61558-1:19, EN 61558-2-15:12

Other data:

The validity of the certificate is limited to: 14.12.2028

15.12.2025

Prague

Mgr. Miroslav Sedláček
Head of Certification Body



* C E R / 1 2 5 0 4 8 7 *

252783-01